

Clean-Up of Organochlorine Pesticides and PCB Extracts Using FLORISIL®



UCT Featured Products

EUFLSA1M6

1000 mg small particle Grade A Florisil® for slower gravity flow

EUFLS1M6

1000 mg regular particle PR Grade Florisil® for more viscous samples

Introduction:

This Florisil® is a magnesium silicate with basic properties. It is used to remove polar interferences from organochlorine pesticide and PCB extracts in hexane prior to chromatographic analysis. Other analytes may potentially be cleaned up using this method provided that adequate performance is demonstrated.

Procedure:

1. Pre-rinse a column with 9 mL of 90:10 hexane/acetone by gravity flow (A low vacuum may be required to start flow).
2. Discard solvent
3. Add a collection tube under the column
4. Add a 2 mL aliquot of the sample extract (in hexane) to the column
5. Collect extract by gravity
6. Add 9 mL of 90:10 hexane/acetone to the column
7. Continue to collect by gravity
8. Gently evaporate the extract to a volume of 1 mL
9. Bring to a final volume of 2 mL with hexane



UCT, LLC • 2731 Bartram Road • Bristol, PA 19007 800.385.3153 • 215.781.9255

www.unitedchem.com Email: methods@unitedchem.com

©UCT, LLC 2025 • All rights reserved

