

2731 Bartram Road Bristol, PA 19007-6893 800.385.3153 www.unitedchem.com

## CERTIFICATE OF ANALYSIS

PRODUCT NAME:ECC18515CTCHEMICAL:Octadecyl endcappedSILICA TYPE:Irregular particles 40-63 micron, 60 AngstromsLOT NUMBER:039232-BH

		Specification Values		<b>Lot Analysis</b>
Test	Method	Min ]	Max	
Average Pore Size	SOP82-4.042	55Å	65Å	60Å
Surface Area (m <sup>2</sup> /g)	SOP82-4.042	470	530	$527 \text{ m}^2/\text{g}$
Pore Volume (cm <sup>3</sup> /g)	SOP82-4.042	0.70	0.85	$0.75 \text{ cm}^3/\text{g}$
Appearance	Visual	White Fir	ne Powder	White Fine Powder
CEC18				
Carbon Analysis by Weight (Organic Loading)	SOP82-4.024	20.50%	22.70%	20.85%
Endcapping	SOP82-4.027	0.00abs (	0.90abs	0.372abs
• Total sorbent weight per tube (500mg)		475	525	Conforms

 All components of this product are individually weighed. The mass of each raw material meets method required tolerances.

The shelf life is only valid if the material is stored in the original unopened container. The specifications and lot analysis are only valid using the listed methods. The test methods and specifications were current when the product was lotted.

## **CERTIFIED BY:**

NAME: D. Shawn Ream

TITLE: QC MANAGER

Certificate issue DATE: 01-30-2019

2731 Bartram Road / Bristol, PA 19007 USA / Phone 215-781-9255 / Fax 215-785-1226

Email: <a href="mailto:info@unitedchem.com">info@unitedchem.com</a> Internet: HTTP://www.unitedchem.com