



2731 Bartram Road
Bristol, PA 19007-6893
800.385.3153
www.unitedchem.com

CERTIFICATE OF ANALYSIS

PRODUCT NAME:

EUPAX1

CHEMICAL:

POLYIMINE BONDED UCT PARTICLE

SILICA TYPE:

Irregular particles 40-63 micron, 60 Angstrom pores

LOT NUMBER:

040058-HO

<u>Test</u>	<u>Method</u>	<u>Specification Values</u>		<u>Lot Analysis</u>
		<u>Min</u>	<u>Max</u>	
Appearance	Visual	White Fine Powder		White Fine Powder
Average Pore Size	SOP82-4.042	55Å	65Å	60Å
Surface Area (m ² /g)	SOP82-4.042	470	530	530 m ² /g
Pore Volume (cm ³ /g)	SOP82-4.042	0.70	0.85	0.75 cm ³ /g
Carbon Analysis by Weight (Organic Loading)	SOP82-4.024	13.50%	15.00%	14.49%
Anion Exchange	SOP82-4.028	0.68	1.05meq/g	0.864meq/g

The shelf life is only valid if the material is stored in the original unopened container.

The specifications and lot analysis are only valid using the listed methods.

The test methods and specifications were current when the product was lotted.

CERTIFIED BY:

NAME: D. Kearns

TITLE: QC ANALYST

Certificate issue DATE: 02-27-2019

2731 Bartram Road / Bristol, PA 19007 USA / Phone 215-781-9255 / Fax 215-785-1226

Email: info@unitedchem.com Internet: [HTTP://www.unitedchem.com](http://www.unitedchem.com)