2731 Bartram Road
Bristol, PA 19007-6893
800.385.3153
www.unitedchem.com

CERTIFICATE OF ANALYSIS

PRODUCT NAME: ECUNIOGXF

CHEMICAL: OCTADECYL ENDCAPPED BONDED SILICA SILICA TYPE: Irregular particles 80-200 micron, 60 Angstrom pores

041763-HM LOT NUMBER:

		Specification Values		Lot Analysis
Test Average Pore Size	Method SOP82-4.042	<u>Min</u> 55Å	<u>Max</u> 65Å	60Å
Surface Area (m ² /g)	SOP82-4.042	470	530	$513\text{m}^2/\text{g}$
Pore Volume (cm ³ /g)	SOP82-4.042	0.70	0.85	$0.75 \text{ cm}^3/\text{g}$
pH	Supplier	6.5	7.5	7.0
XRODH				
Appearance	Visual	White F	Fine Powder	White Fine Powder
Carbon Analysis by Weight (Organic Loading)	SOP82-4.024	20.5%	22.70%	20.50%
Endcapping	SOP82-4.027	0.00abs	0.90abs	0.293abs

The shelf life is only valid if the material is stored in the original unopened container. The specifications and lot analysis are only valid using the listed methods. The test methods and specifications were current when the product was lotted.

CERTIFIED BY: NAME: D. Kearns

TITLE: QC ANALYST

Certificate issue DATE: 04/23/2019

2731 Bartram Road / Bristol, PA 19007 USA / Phone 215-781-9255 / Fax 215-785-1226