



2731 Bartram Road
Bristol, PA 19007-6893
800.385.3153
www.unitedchem.com

CERTIFICATE OF ANALYSIS

PRODUCT NAME: EUMIV50CT
CHEMICAL: Mixture of MgSO₄, NaCl, Sodium citrate dihydrate,
Sodium citrate sesquihydrate
LOT NUMBER: 068194-BO

<u>Test</u>	<u>Method</u>	<u>Specification Values</u>		<u>Lot Analysis</u>
		<u>Min</u>	<u>Max</u>	
MgSO₄ (lot#223010-BO)				
Assay	Ignition	99.50%	100.5%	100.1%
• Total sorbent weight per tube (6000mg)		5700	6300	Conforms
NaCl (lot#128110-BQ)				
Assay		99.00%		100.9%
• Total sorbent weight per tube (1500mg)		1425	1575	Conforms
Sodium citrate dihydrate (lot#129260-BY)				
Assay		99.00%		99.60%
• Total sorbent weight per tube (1500mg)		1425	1575	Conforms
Sodium citrate sesquihydrate (lot#127250-BJ)				
Assay		99.00%		99.90%
• Total sorbent weight per tube (750mg)		712	788	Conforms
• All components of this product are individually weighed. The mass of each raw material meets method required tolerances.				

*The shelf life is only valid if the material is stored in the original unopened container.
The specifications and lot analysis are only valid using the listed methods.
The test methods and specifications were current when the product was lotted.*

CERTIFIED BY:
NAME: L. Bender

TITLE: QC Analyst

Certificate issue DATE: 05/25/2022

2731 Bartram Road / Bristol, PA 19007 USA / Phone 215-781-9255 / Fax 215-785-1226
Email: info@unitedchem.com Internet: [HTTP://www.unitedchem.com](http://www.unitedchem.com)