



2731 Bartram Road
Bristol, PA 19007-6893
800.385.3153
www.unitedchem.com

CERTIFICATE OF ANALYSIS

PRODUCT NAME: XRDAH
CHEMICAL: C8 + Benzyl Sulfonic Acid
SILICA TYPE: Irregular particles 60-200 micron, 60 Angstrom pores
LOT NUMBER: 223101-KQ

<u>Test</u>	<u>Method</u>	<u>Specification Values</u>		<u>Lot Analysis</u>
		<u>Min</u>	<u>Max</u>	
Appearance	Visual	White Fine Powder		Conforms
Average Pore Size	SOP82-4.042	55Å	65Å	60Å
Surface Area (m ² /g)	SOP82-4.042	470	550	528 m ² /g
Pore Volume (cm ³ /g)	SOP82-4.042	0.70	0.85	0.75 cm ³ /g
Carbon Analysis by Weight (Organic Loading)	SOP82-4.024	12.00%	12.80%	12.06%
Cation Exchange	SOP82-4.025	0.065meq/g	0.088meq/g	0.068meq/g
%Recovery - Acid	SOP82-4.011	85.00%	110.0%	Pass
%Recovery - Neutral	SOP82-4.011	85.00%	110.0%	Pass
%Recovery - Base	SOP82-4.011	85.00%	110.0%	Pass

*The shelf life is only valid if the material is stored in the original unopened container.
The specifications and lot analysis are only valid using the listed methods.
The test methods and specifications were current when the product was lotted.*

CERTIFIED BY:

NAME: L. Bender

TITLE: QC Analyst

Certificate issue DATE: 10/21/2022

2731 Bartram Road / Bristol, PA 19007 USA / Phone 215-781-9255 / Fax 215-785-1226
Email: info@unitedchem.com Internet: [HTTP://www.unitedchem.com](http://www.unitedchem.com)