



2731 Bartram Road  
Bristol, PA 19007-6893  
800.385.3153  
www.unitedchem.com

## CERTIFICATE OF ANALYSIS

**PRODUCT NAME:** CUMPSGGC182CT  
**CHEMICAL:** Mixture of MgSO<sub>4</sub>, CUPSA, EEC18, ChloroFiltr  
**SILICA TYPE:** Irregular particles 40-63 micron, 60 Angstroms  
**LOT NUMBER:** 072082-AK

<u>Test</u>	<u>Method</u>	<u>Specification Values</u>		<u>Lot Analysis</u>
		<u>Min</u>	<u>Max</u>	
Average Pore Size	SOP82-4.042	55Å	65Å	60Å
Surface Area (m <sup>2</sup> /g)	SOP82-4.042	470	530	496 m <sup>2</sup> /g
Pore Volume (cm <sup>3</sup> /g)	SOP82-4.042	0.70	0.85	0.73 cm <sup>3</sup> /g
Appearance	Visual	White Fine Powder	White Fine Powder	
<b>MgSO<sub>4</sub></b> (lot#223010-BO)				
Assay	Ignition	99.50%	100.5%	100.1%
• Total sorbent weight per tube (150mg)		143	158	Conforms
<b>CUPSA</b> (lot#128290-GZ)				
Carbon Analysis by Weight (Organic Loading)	SOP82-4.024	10.50%	12.50%	11.57%
Anion Exchange	SOP82-4.028	1.00	1.50	1.116meq/g
• Total sorbent weight per tube (50mg)		48	53	Conforms
<b>EEC18</b> (lot#029130-CC)				
Carbon Analysis by Weight (Organic Loading)	SOP82-4.024	20.50%	22.70%	20.87%
Endcapping	SOP82-4.027	0.00abs	2.00abs	0.733 abs
• Total sorbent weight per tube (50mg)		48	53	Conforms
<b>Chloro Filtr®</b> (lot#20210279)				
• Total sorbent weight per tube (50mg)		Conforms		Conforms
		48	53	Conforms

- All components of this product are individually weighed. The mass of each raw material meets method required tolerances.

*The shelf life is only valid if the material is stored in the original unopened container.*

*The specifications and lot analysis are only valid using the listed methods.*

*The test methods and specifications were current when the product was lotted.*

**CERTIFIED BY:**  
**NAME:** L. Bender

**TITLE:** QC Analyst

**Certificate issue DATE:** 11/15/2022

2731 Bartram Road / Bristol, PA 19007 USA / Phone 215-781-9255 / Fax 215-785-1226  
Email: [info@unitedchem.com](mailto:info@unitedchem.com) Internet: [HTTP://www.unitedchem.com](http://www.unitedchem.com)