



2731 Bartram Road  
Bristol, PA 19007-6893  
800.385.3153  
www.unitedchem.com

## CERTIFICATE OF ANALYSIS

**PRODUCT NAME:** ECUNI549  
**CHEMICAL:** OCTYL ENDCAPPED BONDED SILICA  
**SILICA TYPE:** Irregular particles 40-63 micron, 60 Angstrom pores  
**LOT NUMBER:** 072506-FL

<u>Test</u>	<u>Method</u>	<u>Specification Values</u>		<u>Lot Analysis</u>
		<u>Min</u>	<u>Max</u>	
Average Pore Size	SOP82-4.042	55Å	65Å	60Å
Surface Area (m <sup>2</sup> /g)	SOP82-4.042	470	530	520 m <sup>2</sup> /g
Pore Volume (cm <sup>3</sup> /g)	SOP82-4.042	0.70	0.85	0.78 cm <sup>3</sup> /g
pH	Supplier	6.5	7.5	7.0
<b>CEC08</b>				
Appearance	Visual	White Fine Powder		White Fine Powder
Carbon Analysis by Weight (Organic Loading)	SOP82-4.024	10.0%	12.0%	11.65%
Endcapping	SOP82-4.027	1.0abs	3.0abs	1.380 abs

*The shelf life is only valid if the material is stored in the original unopened container.  
The specifications and lot analysis are only valid using the listed methods.  
The test methods and specifications were current when the product was lotted.*

**CERTIFIED BY:**  
**NAME:** L. Bender

**TITLE:** QC Analyst

**Certificate issue DATE:** 11/30/2022

2731 Bartram Road / Bristol, PA 19007 USA / Phone 215-781-9255 / Fax 215-785-1226  
Email: [info@unitedchem.com](mailto:info@unitedchem.com) Internet: [HTTP://www.unitedchem.com](http://www.unitedchem.com)