

UCT 2731 Bartram Road Bristol, PA 19007-6893 800.385.3153 www.unitedchem.com www.unitedchem.com

Specification Values

CERTIFICATE OF ANALYSIS

PRODUCT NAME: CHEMICAL: SILICA TYPE: LOT NUMBER:

EC525 OCTADECYL UNENDCAPPED BONDED SILICA Irregular particles 40-63 micron, 60 Angstroms 074428-DP

Lot Analysis

<u>Test</u> Appearance	<u>Method</u> Visual	<u>Min</u> <u>Max</u> White Fine Powder	White Fine Powder
Average Pore Size	SOP82-4.042	55Å 65Å	60Å
Surface Area (m^2/g)	SOP82-4.042	470 530	504 m ² /g
Pore Volume (cm^3/g)	SOP82-4.042	0.62 1.02	$0.74 \text{ cm}^{3}/\text{g}$
Carbon Analysis by Weight	SOP82-4.024	15.50% 17.50%	17.13%
(Organic Loading)			

The shelf life is only valid if the material is stored in the original unopened container. The specifications and lot analysis are only valid using the listed methods. The test methods and specifications were current when the product was lotted.

CERTIFIED BY: NAME: D. Kearns

TITLE: QC Analyst

Certificate issue DATE: 03/02/2023

2731 Bartram Road / Bristol, PA 19007 USA / Phone 215-781-9255 / Fax 215-785-1226 Email: <u>info@unitedchem.com</u> Internet: HTTP://www.unitedchem.com