



2731 Bartram Road
Bristol, PA 19007-6893
800.385.3153
www.unitedchem.com

CERTIFICATE OF ANALYSIS

PRODUCT NAME: ECC1850CT
CHEMICAL: OCTADECYL ENDCAPPED BONDED SILICA (CEC18)
SILICA TYPE: Irregular particles 40-63 micron, 60 Angstroms
LOT NUMBER: 075590-DE

		<u>Specification Values</u>		<u>Lot Analysis</u>
<u>Test</u>	<u>Method</u>	<u>Min</u>	<u>Max</u>	
Average Pore Size	SOP82-4.042	55Å	65Å	60Å
Surface Area (m ² /g)	SOP82-4.042	470	530	503 m ² /g
Pore Volume (cm ³ /g)	SOP82-4.042	0.70	0.85	0.70 cm ³ /g
Appearance	Visual	White Fine Powder		White Fine Powder
CEC18				
Carbon Analysis by Weight (Organic Loading)	SOP82-4.024	20.50%	22.70%	21.10%
Endcapping	SOP82-4.027	0.00abs	0.90abs	0.742abs
● Total sorbent weight per tube (500mg)		475	525	Conforms

- All components of this product are individually weighed. The mass of each raw material meets method required tolerances.

*The shelf life is only valid if the material is stored in the original unopened container.
The specifications and lot analysis are only valid using the listed methods.
The test methods and specifications were current when the product was lotted.*

CERTIFIED BY:
NAME: D. Kearns

TITLE: QC Analyst

Certificate issue DATE: 04/26/2023

2731 Bartram Road / Bristol, PA 19007 USA / Phone 215-781-9255 / Fax 215-785-1226
Email: info@unitedchem.com Internet: [HTTP://www.unitedchem.com](http://www.unitedchem.com)