



2731 Bartram Road
Bristol, PA 19007-6893
800.385.3153
www.unitedchem.com

CERTIFICATE OF ANALYSIS

PRODUCT NAME:
LOT NUMBER:

CUCARB
322050-EX

Test

Appearance:
Mesh Size:
Surface Area (m²/g)
GC/MS extraction:

Lot Analysis

Black nonporous, graphitized carbon
120/400
133 m²/g
Pass

*The shelf life for this product is 5 years from the date of shipment.
The shelf life is only valid if the material is stored in the original unopened container.
The test methods and specifications were current when the product was lotted.*

CERTIFIED BY:
NAME: D. Kearns

TITLE: QC Analyst

Certificate issue DATE: 05/09/2023

2731 Bartram Road / Bristol, PA 19007 USA / Phone 215-781-9255 / Fax 215-785-1226
Email: info@unitedchem.com Internet: [HTTP://www.unitedchem.com](http://www.unitedchem.com)