

UCT 2731 Bartram Road Bristol, PA 19007-6893 800.385.3153 www.unitedchem.com www.unitedchem.com

CERTIFICATE OF ANALYSIS

PRODUCT NAME: CHEMICAL: SILICA TYPE: LOT NUMBER:

ECMMCCNAX215CT Mixture of MgSO4, CUNAX1 Irregular particles 40-63 micron, 60 Angstroms 078688-TA

Creation Values I of Analysis

		<u>Specific</u>	cation values	Lot Analysis
Test Average Pore Size Surface Area (m ² /g) Pore Volume (cm ³ /g) Appearance	<u>Method</u> SOP82-4.042 SOP82-4.042 SOP82-4.042 Visual	<u>Min</u> 55Å 470 0.70 White F	Max 65Å 530 0.85 Fine Powder	60Å 496 m ² /g 0.73 cm ³ /g White Fine Powder
MgSO4 (lot#226040-BQ) Assay • Total sorbent weight per tube (1200mg)	Ignition	99.50% 1140	100.5% 1260	100.1% Conforms
CUNAX1 (lot#129110-ON) Carbon Analysis by Weight (Organic Loading)	SOP82-4.024	6.15%	7.15%	6.83%
Anion ExchangeTotal sorbent weight per tube (200mg)	SOP82-4.028	0.265 190	0.340 210	0.310meq/g Conforms

• All components of this product are individually weighed. The mass of each raw material meets method required tolerances.

The shelf life is only valid if the material is stored in the original unopened container. The specifications and lot analysis are only valid using the listed methods. The test methods and specifications were current when the product was lotted.

CERTIFIED BY: NAME: L. Bender

TITLE: QC Analyst

Certificate issue DATE: 09/18/2023

2731 Bartram Road / Bristol, PA 19007 USA / Phone 215-781-9255 / Fax 215-785-1226 Email: info@unitedchem.com Internet: HTTP://www.unitedchem.com