

UCT 2731 Bartram Road Bristol, PA 19007-6893 800.385.3153 www.upitedoborg.com www.unitedchem.com

CERTIFICATE OF ANALYSIS

PRODUCT NAME: CHEMICAL: SILICA TYPE: LOT NUMBER:

ECUNI525 OCTADECYL UNENDCAPPED BONDED SILICA Irregular particles 40-63 micron, 60 Angstroms 079398-DR

Creation Values I of Analysis

		Specification Values		Lot Analysis
Test	Method	<u>Min</u>	<u>Max</u> 65Å	
Average Pore Size	SOP82-4.042	55Å	65Å	60Å
Surface Area (m ² /g)	SOP82-4.042	470	530	504m²/g
Pore Volume (cm^3/g)	SOP82-4.042	0.62	1.02	$0.74 \text{ cm}^{3}/\text{g}$
EU525				
Appearance	Visual	White Fine Powder		White Fine Powder
Carbon Analysis by Weight (Organic Loading)	SOP82-4.024	16.00%	18.00%	16.97%

The shelf life is only valid if the material is stored in the original unopened container. The specifications and lot analysis are only valid using the listed methods. The test methods and specifications were current when the product was lotted.

CERTIFIED BY: NAME: D. Kearns

TITLE: QC Analyst

Certificate issue DATE: 10/27/2023

2731 Bartram Road / Bristol, PA 19007 USA / Phone 215-781-9255 / Fax 215-785-1226 Email: info@unitedchem.com Internet: HTTP://www.unitedchem.com